





**CERTIFICATE OF ANALYSIS**

**03/18/2008**

**NEW YORK DAILY NEWS**

**1892 EAST CHESTER RD**

**BRONX, NY 10461**

**CONTACT: BILL EGBERT**

**CUSTOMER ID:** P.S. 86 QUEENS  
**MATRIX:** SOLID  
**DATE RECEIVED:** 03/06/2008 **TIME:** 10:00  
**SAMPLED BY:** J. WISLOSKI  
**CUSTOMER PO:** N/A

**NEA ID:** AL03559 **NEA LRF:** 08030034-01  
**DATE SAMPLED:** 02/13/2008 **TIME:** 08:15  
**PROJECT:** P.S. 86 QUEENS  
**LOCATION:** JAMAICA QUEENS, NY  
**LAB ELAP#:** 11078

<b>PARAMETER PERFORMED</b>	<b>RESULTS</b>	<b>PQL</b>	<b>UNITS</b>	<b>DATE ANALYZED</b>	<b>FLAGS</b>
<b>SW-846 8082 (PCB)</b>					
Aroclor 1016	ND	13.7	ug/g	03/14/2008	U
Aroclor 1221	ND	13.7	ug/g	03/14/2008	U
Aroclor 1232	ND	13.7	ug/g	03/14/2008	U
Aroclor 1242	ND	13.7	ug/g	03/14/2008	U
Aroclor 1248	ND	13.7	ug/g	03/14/2008	U
Aroclor 1254	357	13.7	ug/g	03/14/2008	AF
Aroclor 1260	161	13.7	ug/g	03/14/2008	PG
Total PCB Amount > Reporting Limit	518				

Notes: ND (Not Detected). Denotes analyte not detected at a concentration greater than the PQL.

PQL (Practical Quantitation Limit). Denotes lowest analyte concentration reportable for the sample.

AF-Aroclor 1254 is being reported as the best Aroclor match. The sample exhibits an altered PCB pattern.

PG-Aroclor 1260 is being used to report an altered PCB pattern exhibited by the sample. Actual Aroclor 1260 is not present in the sample, but is reported to more accurately quantify PCB present in sample that has undergone environmental alteration.

**Note (1.): Sample was received past method specified extraction hold time (hold time=14 days from date sampled).**

**Note (2.): Sample exhibited an Aroclor 1268 pattern. Aroclor 1268 was quantified using an Aroclor 1260 standard which is the most similar Aroclor to A1268 which is routinely calibrated.**

AUTHORIZED SIGNATURE:

William A. Kotas  
Quality Assurance Officer

Robert E. Wagner  
Laboratory Director